

**ANALYTICAL LABORATORY REPORT (SAMPLE)**

WE TREAT CLIENTS BETTER!

PLM by Point Count (400)\*

Client: Your Company  
Address: your address  
Your City

Attention: Project Manager  
Project: Your project location

Client Job #: JOB1  
Laboratory Batch #: 200677777  
Date Received: 1/1/2006  
Samples Received: 1  
Date Analyzed: 1/1/2006  
Samples Analyzed: 1  
Client Project #: N/A

**Sample Requested for Point Count****RC-1****Previous Analytical Information**

Previously Analyzed by: Laboratory  
Previous Batch #: 200588898  
Previous Lab ID: 2005889003  
Previous Description: White compacted powdery material with paint  
Layer that contained asbestos: 1  
Asbestos Type Found: Chrysotile  
Asbestos Percentage Found: 2%

**Point Count Analytical Procedures****New Lab ID: 2006777777**

	<b>Asbestos Points</b>	<b>Non-Asbestos Points</b>	<b>Total Points Counted</b>
Slide 1	1	49	50
Slide 2	0	50	50
Slide 3	1	49	50
Slide 4	0	50	50
Slide 5	1	49	50
Slide 6	0	50	50
Slide 7	0	50	50
Slide 8	0	50	50
Total	3	397	400

**Point Count Summary Results**

Type of Asbestos: Chrysotile  
Percentage of Asbestos: **0.75%**

Analyst Comments: There's no change in sample description nor in layer identification.

\* 600 points and 1000 points also available