

3/16/2006

Fanyao Zhang
Seattle Asbestos Test, LLC
19711 Scriber Lake Road
Suite D
Lynnwood, WA 98036

Lab ID# 181559

Dear Fanyao,

Please find your laboratory's Industrial Hygiene Proficiency Analytical Testing (IHPAT) results for **Round 164**.
IHPAT **Round 165** sample kits will be mailed to laboratories around April 1, 2006. Your laboratory's data will be due by 11:59pm EST on May 1, 2006. The analytes for round 165 are:

- 1 **Metals – cadmium, lead, zinc**
- 2 **Organics – chloroform(CFM), 1,2 dichloroethane(DCE), trichloroethylene(TCE)**
- 3 **Asbestos – chrysotile**
- 4 **Silica – calcite**

Please handle, store and analyze your laboratory's PAT samples in the same manner as routine client samples. To submit your laboratory's data, please visit the Proficiency Analytical Testing (PAT) page and click on the PAT Data Entry Portal:

<http://www.aiha.org/Content/LQAP/PT/pt.htm>

Your laboratory's password needed to access the PAT Data Entry Portal is provided in the upper right hand corner (next to your lab ID#) of the address label on the results submission form included with your PAT samples.


Please note: Print and save the confirmation page after submitting data via the AIHA PAT Data Entry Portal.

The AIHA Laboratory Quality Assurance Programs, Policies and Application for AIHA accreditation are available on-line.

<http://www.aiha.org/Content/LQAP/documents/accredpolicymods.htm>

Note: The Policies for 2005 comply with ISO 17025.

I encourage you to contact me with any feedback, questions or if you wish to contest your results at (703) 846-0797.
Sincerely,



Natasha Sekitoleko
PAT Data Specialist

Industrial Hygiene Proficiency Analytical Testing Results

This document contains three sub-reports relating to IHPAT Round 164. The first report contains your laboratory's results listed per contaminant, per sample. The second report contains your current and 2 previous test round performance respectively (where applicable), and the final report contains summary results for all laboratories for IHPAT round 164.

Testing Results for IHPAT Round 164

This part of the report contains your laboratory's results listed per contaminant, per sample.

Contaminant	Units	#	Result	Mean	Lower Limit	Upper Limit	z-Score	Rati
Asbestos / Fibers (ASB)	f/mm2	1	156.5000	170.3719	83.4822	287.9284	-0.4	A
	f/mm2	2	389.9000	405.5106	210.6648	663.5926	-0.2	A
	f/mm2	3	308.1000	321.5275	171.6856	517.9945	-0.2	A
	f/mm2	4	101.9000	109.0086	53.4142	184.2246	-0.3	A

**Please note: A – Acceptable; U - Unacceptable